



**SCOTTSDALE POLICE DEPARTMENT  
CRIME LABORATORY  
FORENSIC BLOOD ALCOHOL FACE SHEET**

DATE: 073014 SEQUENCE FILE NAME: 28Jul14VPV

Pipettor: *Hamilton MicroLab*

Gas Chromatograph: *Perkin Elmer*

☐ MD91JC4962 ☒ MD91JF5016  
☒ 650N9042002 ☐ 650N9042003

INTERNAL STANDARD

N-Propanol 0.015% (w/v)

Lot # 050214VPV-B

CALIBRATION STANDARDS

Vial #1 0.020 Calibrator

Lot # FN09031301

Vial #2 0.100 Calibrator

Lot # FN050312-01

Vial #3 0.200 Calibrator

Lot # FN032712-01

Vial #4 0.400 Calibrator

Lot # FN012712-01

Plot Analysis:

R<sup>2</sup>= 0.999974

CONTROLS

Vial #5 & 103ISTD BLANK

Vial #6 MIX STANDARD

Lot # 011111VPV

| <u>Vial #</u> | <u>Expected [Ethanol] g/dL</u> | <u>Measured [Ethanol] g/dL</u> | <u>Lot #</u>      |
|---------------|--------------------------------|--------------------------------|-------------------|
| <u>07</u>     | <u>0.400</u>                   | <u>0.397</u>                   | <u>05012012-C</u> |
| <u>08</u>     | <u>0.040</u>                   | <u>0.040</u>                   | <u>30112011-B</u> |
| <u>53</u>     | <u>0.202</u>                   | <u>0.201</u>                   | <u>1212091</u>    |
| <u>54</u>     | <u>0.202</u>                   | <u>0.201</u>                   | <u>1212091</u>    |
| <u>55</u>     | <u>0.150</u>                   | <u>0.149</u>                   | <u>11012012-C</u> |
| <u>56</u>     | <u>0.080</u>                   | <u>0.079</u>                   | <u>14112011-A</u> |
| <u>104</u>    | <u>0.400</u>                   | <u>0.398</u>                   | <u>05012012-C</u> |
| <u>105</u>    | <u>0.040</u>                   | <u>0.040</u>                   | <u>30112011-B</u> |

Notes: 45 subject samples

Chain of Custody – 072814

Exam Notes – 072814 (Samples 1-22) & 072914 (Samples 23-45)

Analysis – 072914; All testing proceeded as expected.

VALIDATED RESULTS: ✓ B1254 073014

Analyst

A B1255 8/4/14

Technical Reviewer

Original Adoption Date: February 25th, 2010 revised 7-21-11

Version: F-TOX-BA\_001

Version Effective Date: 7-21-11

Issuing Authority: Kris Whitman, QAM

Fit Analysis Output For Method File: C:\TCWMETHODS\METHOD2002A.MTH

Reprocess Number : spdmscpe01: 2018

Component Name : Ethanol

Date : 7/29/2014 4:00:55 PM

## Curve Parameters:


Curve #1 : 1st Order

Weighting Factor = 1 (No Weighting)

R-Squared = 0.999974

Calibration Curve :  $Y = (-0.004558) + (3.552470) X$ 

Ethanol




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.020 CAL (FN09031301)


Vial #: 1



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.080 | 19587  | Ethanol    |
| 1.773 | 282787 | n-Propanol |

Sample: 0.020 CAL (FN09031301)

Vial #: 1




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.703 | 16246  | Ethanol    |
| 3.319 | 249058 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.100 CAL (FN050312-01)


Vial #: 2



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.076 | 97160  | Ethanol    |
| 1.769 | 277877 | n-Propanol |

Sample: 0.100 CAL (FN050312-01)

Vial #: 2




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 86206  | Ethanol    |
| 3.315 | 244455 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: 0.200 CAL (FN032712-01)

Vial #: 3



Sample: 0.200 CAL (FN032712-01)

Vial #: 3




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.400 CAL (FN012712-01)


Vial #: 4



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.077 | 393341 | Ethanol    |
| 1.771 | 277311 | n-Propanol |

Sample: 0.400 CAL (FN012712-01)

Vial #: 4




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 355944 | Ethanol    |
| 3.317 | 242496 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: IS Blank (050214VPV-B)


Vial #: 5



| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.770 | 272194 | n-Propanol |       |

Sample: IS Blank (050214VPV-B)

Vial #: 5




| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 3.315 | 237051 | n-Propanol |       |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: MIX (011111VPV)


Vial #: 6



| Rt    | Area   | Peak         | g/100  |
|-------|--------|--------------|--------|
| 0.860 | 38950  | Methanol     | 0.0803 |
| 0.957 | 118225 | Acetaldehyde |        |
| 1.076 | 76009  | Ethanol      |        |
| 1.313 | 146159 | Isopropanol  |        |
| 1.613 | 93940  | Acetone      |        |
| 1.770 | 270521 | n-Propanol   |        |

Sample: MIX (011111VPV)

Vial #: 6




| Rt    | Area   | Peak         |
|-------|--------|--------------|
| 1.066 | 109447 | Acetaldehyde |
| 1.260 | 34466  | Methanol     |
| 1.699 | 67023  | Ethanol      |
| 1.837 | 85168  | Acetone      |
| 2.021 | 130977 | Isopropanol  |
| 3.316 | 234742 | n-Propanol   |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.400 QC (05012012-C)


Vial #: 7



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 396384 | Ethanol    | 0.3973 |
| 1.771 | 281693 | n-Propanol |        |

Sample: 0.400 QC (05012012-C)

Vial #: 7




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 358754 | Ethanol    |
| 3.317 | 245415 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.040 QC (30112011-B)


Vial #: 8



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 38523  | Ethanol    | 0.0401 |
| 1.771 | 279071 | n-Propanol |        |

Sample: 0.040 QC (30112011-B)

Vial #: 8




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 33396  | Ethanol    |
| 3.318 | 242634 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.202 WB QC (1212091)


Vial #: 53



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 196150 | Ethanol    | 0.2010 |
| 1.770 | 276458 | n-Propanol |        |

Sample: 0.202 WB QC (1212091)

Vial #: 53




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 174929 | Ethanol    |
| 3.319 | 240204 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.202 WB QC (1212091)


Vial #: 54



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 197788 | Ethanol    | 0.2016 |
| 1.770 | 277885 | n-Propanol |        |

Sample: 0.202 WB QC (1212091)

Vial #: 54




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 175860 | Ethanol    |
| 3.318 | 240250 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.150 QC (11012012-C)


Vial #: 55



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 145886 | Ethanol    | 0.1491 |
| 1.770 | 277734 | n-Propanol |        |

Sample: 0.150 QC (11012012-C)

Vial #: 55




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 129130 | Ethanol    |
| 3.317 | 241063 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.080 QC (14112011-A)


Vial #: 56



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 77257  | Ethanol    | 0.0793 |
| 1.770 | 278453 | n-Propanol |        |

Sample: 0.080 QC (14112011-A)

Vial #: 56




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 67652  | Ethanol    |
| 3.318 | 241664 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: IS Blank (050214VPV-B)


Vial #: 103



| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.769 | 280881 | n-Propanol |       |

Sample: IS Blank (050214VPV-B)

Vial #: 103




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 3.317 | 241383 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.400 QC (05012012-C)


Vial #: 104



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.079 | 399808 | Ethanol    | 0.3981 |
| 1.773 | 283606 | n-Propanol |        |

Sample: 0.400 QC (05012012-C)

Vial #: 104




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 360998 | Ethanol    |
| 3.321 | 245661 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: 0.040 QC (30112011-B)


Vial #: 105



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 39283  | Ethanol    | 0.0402 |
| 1.770 | 283456 | n-Propanol |        |

Sample: 0.040 QC (30112011-B)

Vial #: 105



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 33411  | Ethanol    |
| 3.318 | 245593 | n-Propanol |

TotalChrom Sequence File C:\tcw\sequences\VPV\28Jul14VPV.seq

Printed by : manager on: 7/29/2014 2:02:03 PM

Created by : manager on: 7/29/2014 2:01:57 PM

Edited by : manager on: 7/29/2014 2:01:57 PM

Number of Times Edited : 0

Description:

**Sequence File Header Information:**

Number of Rows : 106  
 Instrument Type : PE AutoSystem GC  
 Injection Type : SINGLE  
 Raw tokens channel A :  
 Result tokens channel A :  
 Modified tokens channel A :  
 Raw tokens channel B :  
 Result tokens channel B :  
 Modified tokens channel B :

**Sequence Sample Descriptions - Channel A**

| Row | Type        | Name                    | Number |
|-----|-------------|-------------------------|--------|
| 1   | Cal:Replace | 0.020 CAL (FN09031301)  | 001    |
| 2   | Cal:Replace | 0.100 CAL (FN050312-01) | 002    |
| 3   | Cal:Replace | 0.200 CAL (FN032712-01) | 003    |
| 4   | Cal:Replace | 0.400 CAL (FN012712-01) | 004    |
| 5   | Blank       | IS Blank (050214VPV-B)  | 005    |
| 6   | Ctrl Sample | MIX (011111VPV)         | 006    |
| 7   | Ctrl Sample | 0.400 QC (05012012-C)   | 007    |
| 8   | Ctrl Sample | 0.040 QC (30112011-B)   | 008    |
| 9   | Sample      |                         | 009    |
| 10  | Sample      |                         | 010    |
| 11  | Sample      |                         | 011    |
| 12  | Sample      |                         | 012    |
| 13  | Sample      |                         | 013    |
| 14  | Sample      |                         | 014    |
| 15  | Sample      |                         | 015    |
| 16  | Sample      |                         | 016    |
| 17  | Sample      |                         | 017    |
| 18  | Sample      |                         | 018    |
| 19  | Sample      |                         | 019    |
| 20  | Sample      |                         | 020    |
| 21  | Sample      |                         | 021    |
| 22  | Sample      |                         | 022    |
| 23  | Sample      |                         | 023    |
| 24  | Sample      |                         | 024    |
| 25  | Sample      |                         | 025    |
| 26  | Sample      |                         | 026    |
| 27  | Sample      |                         | 027    |
| 28  | Sample      |                         | 028    |
| 29  | Sample      |                         | 029    |
| 30  | Sample      |                         | 030    |
| 31  | Sample      |                         | 031    |
| 32  | Sample      |                         | 032    |
| 33  | Sample      |                         | 033    |
| 34  | Sample      |                         | 034    |
| 35  | Sample      |                         | 035    |
| 36  | Sample      |                         | 036    |
| 37  | Sample      |                         | 037    |
| 38  | Sample      |                         | 038    |
| 39  | Sample      |                         | 039    |
| 40  | Sample      |                         | 040    |
| 41  | Sample      |                         | 041    |
| 42  | Sample      |                         | 042    |
| 43  | Sample      |                         | 043    |
| 44  | Sample      |                         | 044    |
| 45  | Sample      |                         | 045    |
| 46  | Sample      |                         | 046    |
| 47  | Sample      |                         | 047    |
| 48  | Sample      |                         | 048    |
| 49  | Sample      |                         | 049    |
| 50  | Sample      |                         | 050    |
| 51  | Sample      |                         | 051    |
| 52  | Sample      |                         | 052    |
| 53  | Ctrl Sample | 0.202 WB QC (1212091)   | 053    |
| 54  | Ctrl Sample | 0.202 WB QC (1212091)   | 054    |
| 55  | Ctrl Sample | 0.150 QC (11012012-C)   | 055    |
| 56  | Ctrl Sample | 0.080 QC (14112011-A)   | 056    |
| 57  | Sample      |                         | 057    |

7/29/2014 2:02:03 PM Sequence: C:\tcw\sequences\VPV\28Jul14VPV.seq

**Sequence Sample Descriptions - Channel A**

| Row | Type        | Name                   | Number |
|-----|-------------|------------------------|--------|
| 58  | Sample      |                        | 058    |
| 59  | Sample      |                        | 059    |
| 60  | Sample      |                        | 060    |
| 61  | Sample      |                        | 061    |
| 62  | Sample      |                        | 062    |
| 63  | Sample      |                        | 063    |
| 64  | Sample      |                        | 064    |
| 65  | Sample      |                        | 065    |
| 66  | Sample      |                        | 066    |
| 67  | Sample      |                        | 067    |
| 68  | Sample      |                        | 068    |
| 69  | Sample      |                        | 069    |
| 70  | Sample      |                        | 070    |
| 71  | Sample      |                        | 071    |
| 72  | Sample      |                        | 072    |
| 73  | Sample      |                        | 073    |
| 74  | Sample      |                        | 074    |
| 75  | Sample      |                        | 075    |
| 76  | Sample      |                        | 076    |
| 77  | Sample      |                        | 077    |
| 78  | Sample      |                        | 078    |
| 79  | Sample      |                        | 079    |
| 80  | Sample      |                        | 080    |
| 81  | Sample      |                        | 081    |
| 82  | Sample      |                        | 082    |
| 83  | Sample      |                        | 083    |
| 84  | Sample      |                        | 084    |
| 85  | Sample      |                        | 085    |
| 86  | Sample      |                        | 086    |
| 87  | Sample      |                        | 087    |
| 88  | Sample      |                        | 088    |
| 89  | Sample      |                        | 089    |
| 90  | Sample      |                        | 090    |
| 91  | Sample      |                        | 091    |
| 92  | Sample      |                        | 092    |
| 93  | Sample      |                        | 093    |
| 94  | Sample      |                        | 094    |
| 95  | Sample      |                        | 095    |
| 96  | Sample      |                        | 096    |
| 97  | Sample      |                        | 097    |
| 98  | Sample      |                        | 098    |
| 99  | Sample      |                        | 099    |
| 100 | Sample      |                        | 100    |
| 101 | Sample      |                        | 101    |
| 102 | Sample      |                        | 102    |
| 103 | Blank       | IS Blank (050214VPV-B) | 103    |
| 104 | Ctrl Sample | 0.400 QC (05012012-C)  | 104    |
| 105 | Ctrl Sample | 0.040 QC (30112011-B)  | 105    |
| 106 | Sample      | END OF SEQUENCE        | 106    |

# **Scottsdale P.D. Crime Lab** **Summary of Cases for Ethanol**

Acquired: 07/29/2014

Analyst: VPV

Sequence: C:\tcw\sequences\VPV\28Jul14VPV.seq

| File        | Sample | Result | Result | % Diff. |
|-------------|--------|--------|--------|---------|
| 29_A009.rst |        | 0.1818 | 0.1813 | 0.28    |
| 29_A011.rst |        | 0.1521 | 0.1520 | 0.10    |
| 29_A013.rst |        | 0.1327 | 0.1323 | 0.33    |
| 29_A015.rst |        | 0.0828 | 0.0829 | 0.21    |
| 29_A017.rst |        | 0.1374 | 0.1379 | 0.33    |
| 29_A019.rst |        | 0.1583 | 0.1588 | 0.29    |
| 29_A021.rst |        | 0.1551 | 0.1553 | 0.14    |
| 29_A023.rst |        | 0.0973 | 0.0978 | 0.47    |
| 29_A025.rst |        | 0.1183 | 0.1178 | 0.42    |
| 29_A027.rst |        | 0.1595 | 0.1598 | 0.22    |
| 29_A029.rst |        | 0.0437 | 0.0439 | 0.26    |
| 29_A031.rst |        | 0.0822 | 0.0822 | 0.09    |
| 29_A033.rst |        | 0.0979 | 0.0978 | 0.06    |
| 29_A035.rst |        | 0.2747 | 0.2738 | 0.32    |
| 29_A037.rst |        | 0.1278 | 0.1280 | 0.13    |
| 29_A039.rst |        | 0.1578 | 0.1578 | 0.02    |
| 29_A041.rst |        | 0.1126 | 0.1123 | 0.25    |
| 29_A043.rst |        | 0.1614 | 0.1611 | 0.15    |
| 29_A045.rst |        | 0.0274 | 0.0275 | 0.60    |
| 29_A047.rst |        | 0.0145 | 0.0149 | 2.53    |
| 29_A049.rst |        | 0.1201 | 0.1199 | 0.16    |
| 29_A051.rst |        | 0.1831 | 0.1831 | 0.01    |
| 29_A057.rst |        | 0.1258 | 0.1280 | 1.79    |
| 29_A059.rst |        | 0.1213 | 0.1215 | 0.18    |
| 29_A061.rst |        | 0.1882 | 0.1881 | 0.05    |
| 29_A063.rst |        | 0.1229 | 0.1226 | 0.26    |
| 29_A065.rst |        | 0.1872 | 0.1870 | 0.10    |
| 29_A067.rst |        | 0.1461 | 0.1456 | 0.31    |
| 29_A069.rst |        | 0.2582 | 0.2558 | 0.91    |
| 29_A071.rst |        | 0.2619 | 0.2611 | 0.28    |
| 29_A073.rst |        | 0.0105 | 0.0104 | 1.26    |
| 29_A075.rst |        | 0.0995 | 0.0990 | 0.49    |
| 29_A077.rst |        | 0.1068 | 0.1067 | 0.05    |
| 29_A079.rst |        | 0.1516 | 0.1510 | 0.41    |
| 29_A081.rst |        | 0.1586 | 0.1587 | 0.05    |
| 29_A083.rst |        | 0.1078 | 0.1073 | 0.46    |
| 29_A085.rst |        | 0.0811 | 0.0814 | 0.37    |
| 29_A087.rst |        | 0.0926 | 0.0923 | 0.34    |
| 29_A089.rst |        | 0.1310 | 0.1307 | 0.20    |
| 29_A091.rst |        | 0.2261 | 0.2258 | 0.10    |
| 29_A093.rst |        | 0.1356 | 0.1361 | 0.36    |
| 29_A095.rst |        | 0.1475 | 0.1474 | 0.08    |
| 30_A097.rst |        | 0.1463 | 0.1467 | 0.31    |
| 30_A099.rst |        | 0.1135 | 0.1137 | 0.22    |

# Scottsdale P.D. Crime Lab

## Summary of Cases for Ethanol

Acquired: 07/29/2014

Analyst: VPV

Sequence: C:\tcw\sequences\VPV28Jul14\VPV.seq \/\


| File        | Sample          | Result | Result | % Diff. |
|-------------|-----------------|--------|--------|---------|
| 30_A101.rst |                 | 0.2250 | 0.2252 | 0.09    |
| 30_A106.rst | END OF SEQUENCE | 0      |        |         |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 9



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 175501 | Ethanol    | 0.1818 |
| 1.771 | 273607 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 9




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 158515 | Ethanol    |
| 3.318 | 236843 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14 [REDACTED]


Vial #: 10



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 182078 | Ethanol    | 0.1813 |
| 1.771 | 284654 | n-Propanol |        |

Sample: L14 [REDACTED]

Vial #: 10




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 164501 | Ethanol    |
| 3.318 | 245679 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 11



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 152216 | Ethanol    | 0.1521 |
| 1.769 | 283970 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 11




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 136583 | Ethanol    |
| 3.317 | 245471 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 12



Sample: L14- [REDACTED]

Vial #: 12




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 134970 | Ethanol    |
| 3.316 | 243261 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 13



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 132463 | Ethanol    | 0.1327 |
| 1.769 | 283630 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 13




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 118887 | Ethanol    |
| 3.316 | 244812 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 14



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 135054 | Ethanol    | 0.1323 |
| 1.769 | 290155 | n-Propanol |        |

Sample: L14-

Vial #: 14




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 121145 | Ethanol    |
| 3.316 | 251689 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 15



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 83228  | Ethanol    | 0.0828 |
| 1.770 | 287350 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 15




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 74509  | Ethanol    |
| 3.317 | 248567 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 16



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 81742  | Ethanol    | 0.0829 |
| 1.770 | 281632 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 16




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 73095  | Ethanol    |
| 3.317 | 243745 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 17



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 138304 | Ethanol    | 0.1374 |
| 1.770 | 285819 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 17




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 124756 | Ethanol    |
| 3.317 | 245140 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 18



Sample: L14- [REDACTED]

Vial #: 18




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 19



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 157803 | Ethanol    | 0.1583 |
| 1.770 | 282779 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 19




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 141739 | Ethanol    |
| 3.318 | 244534 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 20



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 157852 | Ethanol    | 0.1588 |
| 1.770 | 282053 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 20




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 141463 | Ethanol    |
| 3.316 | 243208 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 21



Sample: L14- [REDACTED]

Vial #: 21




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 138216 | Ethanol    |
| 3.318 | 243169 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 22



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 153049 | Ethanol    | 0.1553 |
| 1.771 | 279554 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 22




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 137739 | Ethanol    |
| 3.317 | 242552 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 23



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 97036  | Ethanol    | 0.0973 |
| 1.771 | 284336 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 23




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 87366  | Ethanol    |
| 3.318 | 246196 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 24



Sample: L14- [REDACTED]

Vial #: 24




| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.699 | 86313  | Ethanol    |       |
| 3.317 | 244788 | n-Propanol |       |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 25



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 118797 | Ethanol    | 0.1183 |
| 1.770 | 285772 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 25




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 106252 | Ethanol    |
| 3.317 | 247912 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 26



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 119527 | Ethanol    | 0.1178 |
| 1.771 | 288736 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 26




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 107148 | Ethanol    |
| 3.318 | 249444 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 27



Sample: L14- [REDACTED]

Vial #: 27




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 145505 | Ethanol    |
| 3.320 | 247893 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14 [REDACTED]


Vial #: 28



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 159994 | Ethanol    | 0.1598 |
| 1.771 | 284031 | n-Propanol |        |

Sample: L14 [REDACTED]

Vial #: 28




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 143720 | Ethanol    |
| 3.318 | 244565 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14 [REDACTED]


Vial #: 29



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 42652  | Ethanol    | 0.0437 |
| 1.771 | 282464 | n-Propanol |        |

Sample: L14 [REDACTED]

Vial #: 29




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 38458  | Ethanol    |
| 3.319 | 243912 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 30



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 42636  | Ethanol    | 0.0439 |
| 1.771 | 281597 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 30




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 38189  | Ethanol    |
| 3.319 | 242083 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 31



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 80959  | Ethanol    | 0.0822 |
| 1.772 | 281568 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 31




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 72459  | Ethanol    |
| 3.319 | 244623 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 32



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 81059  | Ethanol    | 0.0822 |
| 1.772 | 281647 | n-Propanol |        |

Sample: L14-

Vial #: 32




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 72397  | Ethanol    |
| 3.319 | 243144 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 33



Sample: L14- [REDACTED]

Vial #: 33




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 86903  | Ethanol    |
| 3.320 | 244228 | n-Propanol |

## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 34



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 97556  | Ethanol    | 0.0978 |
| 1.772 | 284266 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 34




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 87425  | Ethanol    |
| 3.320 | 244936 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 35



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 277056 | Ethanol    | 0.2747 |
| 1.772 | 285236 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 35




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 250210 | Ethanol    |
| 3.320 | 245837 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 36



Sample: L14- [REDACTED]

Vial #: 36




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 37



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.079 | 129060 | Ethanol    | 0.1278 |
| 1.773 | 287065 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 37




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 115873 | Ethanol    |
| 3.320 | 248290 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 38



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.074 | 128362 | Ethanol    | 0.1280 |
| 1.768 | 285144 | n-Propanol |        |

Sample: L14-

Vial #: 38




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 115541 | Ethanol    |
| 3.315 | 247112 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 39



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 157406 | Ethanol    | 0.1578 |
| 1.613 | 6912   | Acetone    |        |
| 1.769 | 283072 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 39




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 140849 | Ethanol    |
| 3.316 | 243971 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14 [REDACTED]


Vial #: 40



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 159065 | Ethanol    | 0.1578 |
| 1.613 | 6915   | Acetone    |        |
| 1.768 | 285990 | n-Propanol |        |

Sample: L14 [REDACTED]

Vial #: 40




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 142150 | Ethanol    |
| 3.317 | 245786 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 41



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 112111 | Ethanol    | 0.1126 |
| 1.769 | 283448 | n-Propanol |        |

Sample: L14 [REDACTED]

Vial #: 41




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 100286 | Ethanol    |
| 3.317 | 244293 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 42



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 112313 | Ethanol    | 0.1123 |
| 1.769 | 284688 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 42




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 100483 | Ethanol    |
| 3.316 | 246755 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 43



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 160265 | Ethanol    | 0.1614 |
| 1.769 | 281689 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 43




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 144269 | Ethanol    |
| 3.317 | 243575 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 44



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 160539 | Ethanol    | 0.1611 |
| 1.769 | 282591 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 44




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 143966 | Ethanol    |
| 3.317 | 243743 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 45



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 26150  | Ethanol    | 0.0274 |
| 1.769 | 281535 | n-Propanol |        |

Sample: L14-

Vial #: 45




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 20355  | Ethanol    |
| 3.315 | 243440 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 46



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 26596  | Ethanol    | 0.0275 |
| 1.769 | 284531 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 46




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 20468  | Ethanol    |
| 3.317 | 244521 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 47



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 13301  | Ethanol    | 0.0145 |
| 1.769 | 282456 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 47




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 9885   | Ethanol    |
| 3.316 | 243377 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 48



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 13645  | Ethanol    | 0.0149 |
| 1.769 | 281926 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 48




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 9706   | Ethanol    |
| 3.317 | 243088 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 49



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 119277 | Ethanol    | 0.1201 |
| 1.769 | 282577 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 49




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 107655 | Ethanol    |
| 3.317 | 244038 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 50



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 120719 | Ethanol    | 0.1199 |
| 1.769 | 286466 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 50




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 109085 | Ethanol    |
| 3.317 | 246954 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 51



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 184347 | Ethanol    | 0.1831 |
| 1.770 | 285363 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 51



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 166102 | Ethanol    |
| 3.318 | 244982 | n-Propanol |


Case: L14- [REDACTED] User: villena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 52



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 183532 | Ethanol    | 0.1831 |
| 1.770 | 284140 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 52




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 164877 | Ethanol    |
| 3.317 | 245806 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 57



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 126196 | Ethanol    | 0.1258 |
| 1.771 | 285261 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 57




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 114107 | Ethanol    |
| 3.318 | 245807 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 58



Sample: L14- [REDACTED]

Vial #: 58




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 59



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 121517 | Ethanol    | 0.1213 |
| 1.771 | 284864 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 59




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 109206 | Ethanol    |
| 3.317 | 244658 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 60



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 124757 | Ethanol    | 0.1215 |
| 1.771 | 291920 | n-Propanol |        |

Sample: L14-

Vial #: 60




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 112172 | Ethanol    |
| 3.319 | 251176 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 61



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 192135 | Ethanol    | 0.1882 |
| 1.771 | 289214 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 61




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 173918 | Ethanol    |
| 3.320 | 249141 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 62



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 192131 | Ethanol    | 0.1881 |
| 1.771 | 289352 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 62




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 173722 | Ethanol    |
| 3.318 | 250774 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 63



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 122808 | Ethanol    | 0.1229 |
| 1.771 | 284200 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 63



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 110244 | Ethanol    |
| 3.318 | 244657 | n-Propanol |


Case: L14- [REDACTED] User: villena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 64



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 123421 | Ethanol    | 0.1226 |
| 1.772 | 286369 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 64



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 110782 | Ethanol    |
| 3.320 | 249528 | n-Propanol |


Case: L14- [REDACTED] User: vwillena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 65



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 189076 | Ethanol    | 0.1872 |
| 1.768 | 286146 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 65




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 170045 | Ethanol    |
| 3.317 | 246739 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 66



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 187231 | Ethanol    | 0.1870 |
| 1.769 | 283644 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 66



| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.697 | 168583 | Ethanol    |       |
| 3.318 | 245228 | n-Propanol |       |


Case: L14- [REDACTED] User: vwillena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 67



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 146333 | Ethanol    | 0.1461 |
| 1.769 | 284354 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 67




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 131913 | Ethanol    |
| 3.317 | 245516 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 68



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 147487 | Ethanol    | 0.1456 |
| 1.769 | 287494 | n-Propanol |        |

Sample: L14-

Vial #: 68




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 132578 | Ethanol    |
| 3.317 | 248474 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 69



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 257250 | Ethanol    | 0.2582 |
| 1.769 | 281827 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 69




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 231911 | Ethanol    |
| 3.318 | 243369 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 70



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 257862 | Ethanol    | 0.2558 |
| 1.770 | 285090 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 70




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 232654 | Ethanol    |
| 3.318 | 246488 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 71



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 265543 | Ethanol    | 0.2619 |
| 1.769 | 286800 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 71




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 239058 | Ethanol    |
| 3.317 | 247052 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 72



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 263250 | Ethanol    | 0.2611 |
| 1.770 | 285131 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 72




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 237444 | Ethanol    |
| 3.317 | 247123 | n-Propanol |

## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 73



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 9339   | Ethanol    | 0.0105 |
| 1.770 | 283971 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 73




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 3.317 | 246095 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 74



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 9333   | Ethanol    | 0.0104 |
| 1.770 | 287831 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 74




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.702 | 6513   | Ethanol    |
| 3.318 | 248574 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 75



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 98859  | Ethanol    | 0.0995 |
| 1.770 | 283168 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 75




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 89409  | Ethanol    |
| 3.317 | 242407 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 76



Sample: L14- [REDACTED]

Vial #: 76




| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.699 | 89516  | Ethanol    |       |
| 3.317 | 244450 | n-Propanol |       |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 77



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 107498 | Ethanol    | 0.1068 |
| 1.770 | 286678 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 77




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 96638  | Ethanol    |
| 3.318 | 248923 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 78



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 110156 | Ethanol    | 0.1067 |
| 1.771 | 293906 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 78




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 99077  | Ethanol    |
| 3.318 | 253508 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 79



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 151737 | Ethanol    | 0.1516 |
| 1.771 | 283984 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 79




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 136718 | Ethanol    |
| 3.319 | 245115 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 80



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 151529 | Ethanol    | 0.1510 |
| 1.770 | 284771 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 80




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 136833 | Ethanol    |
| 3.318 | 245657 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 81



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 160638 | Ethanol    | 0.1586 |
| 1.771 | 287308 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 81




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.699 | 144628 | Ethanol    |
| 3.318 | 247293 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 82



Sample: L14- [REDACTED]

Vial #: 82




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 83



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 107105 | Ethanol    | 0.1078 |
| 1.771 | 282900 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 83




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 96317  | Ethanol    |
| 3.318 | 244278 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 84



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 107034 | Ethanol    | 0.1073 |
| 1.771 | 284020 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 84




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 96431  | Ethanol    |
| 3.320 | 244427 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 85



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.077 | 81900  | Ethanol    | 0.0811 |
| 1.771 | 288729 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 85




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 74199  | Ethanol    |
| 3.319 | 247491 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Sample: L14- [REDACTED] Instrument: 650N9042002 Analyst: VPV


Vial #: 86



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 80634  | Ethanol    | 0.0814 |
| 1.771 | 283194 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 86




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 72825  | Ethanol    |
| 3.319 | 243520 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 87



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 90816  | Ethanol    | 0.0926 |
| 1.771 | 279804 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 87



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 81451  | Ethanol    |
| 3.320 | 241408 | n-Propanol |


Case: L14- [REDACTED] User: villena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 88



Sample: L14- [REDACTED]

Vial #: 88




| Rt    | Area   | Peak       | g/100 |
|-------|--------|------------|-------|
| 1.701 | 82090  | Ethanol    |       |
| 3.322 | 244429 | n-Propanol |       |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 89



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 132359 | Ethanol    | 0.1310 |
| 1.772 | 287198 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 89




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 118877 | Ethanol    |
| 3.321 | 247735 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 90



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 133248 | Ethanol    | 0.1307 |
| 1.772 | 289721 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 90




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.701 | 119955 | Ethanol    |
| 3.321 | 250176 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 91



Sample: L14- [REDACTED]

Vial #: 91




## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 92



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.078 | 230506 | Ethanol    | 0.2258 |
| 1.772 | 288900 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 92



| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 207755 | Ethanol    |
| 3.321 | 249994 | n-Propanol |


Case: L14- [REDACTED] User: villena 7/30/2014

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 93



Sample: L14- [REDACTED]

Vial #: 93




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV


Sample: L14- [REDACTED]

Vial #: 94



Sample: L14- [REDACTED]

Vial #: 94




# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 95



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.074 | 147458 | Ethanol    | 0.1475 |
| 1.768 | 283822 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 95




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 132345 | Ethanol    |
| 3.315 | 243578 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 96



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 149389 | Ethanol    | 0.1474 |
| 1.768 | 287761 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 96




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 134123 | Ethanol    |
| 3.316 | 247157 | n-Propanol |

## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 97



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 148887 | Ethanol    | 0.1463 |
| 1.769 | 288921 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 97




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 134195 | Ethanol    |
| 3.316 | 249211 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 98



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.075 | 147874 | Ethanol    | 0.1467 |
| 1.769 | 286070 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 98




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.697 | 132813 | Ethanol    |
| 3.316 | 245680 | n-Propanol |

## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14-


Vial #: 99



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 113530 | Ethanol    | 0.1135 |
| 1.770 | 284688 | n-Propanol |        |

Sample: L14-

Vial #: 99




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 102017 | Ethanol    |
| 3.317 | 244279 | n-Propanol |

## Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 100



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 112953 | Ethanol    | 0.1137 |
| 1.769 | 282606 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 100




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 101331 | Ethanol    |
| 3.317 | 243724 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 101



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.076 | 227185 | Ethanol    | 0.2250 |
| 1.770 | 285752 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 101




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.698 | 204146 | Ethanol    |
| 3.317 | 246627 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: L14- [REDACTED]


Vial #: 102



| Rt    | Area   | Peak       | g/100  |
|-------|--------|------------|--------|
| 1.079 | 229489 | Ethanol    | 0.2252 |
| 1.773 | 288382 | n-Propanol |        |

Sample: L14- [REDACTED]

Vial #: 102




| Rt    | Area   | Peak       |
|-------|--------|------------|
| 1.700 | 207233 | Ethanol    |
| 3.320 | 248704 | n-Propanol |

# Scottsdale P. D. Crime Lab: Dual Channel Blood Alcohol Report

Instrument: 650N9042002 Analyst: VPV

Sample: END OF SEQUENCE


Vial#: 106



| Rt | Area | Peak | g/100 |
|----|------|------|-------|
|----|------|------|-------|

Sample: END OF SEQUENCE

Vial#: 106



| Rt | Area | Peak |
|----|------|------|
|----|------|------|